



CERTIFICATE OF ANALYSIS

Honeywell Specialty Chemicals Seelze GmbH

Wunstorferstrasse 40
Seelze, 30926
Telefon: +49 5137 999-0
www.lab-honeywell.com

Brand: Honeywell Fluka™
Product: 242143
Lot No: H051S
Production date: 20.Feb.2017
Expiry Date: 05.Jan.2022

Buffer solution DURACAL pH 4.01 (25oC) 250 mL, certified buffer solution traceable to NIST/PTB reference materials, DKD certified

Parameter	Specification	Units	Result	Units
appearance (color)	complying		complying	
pH (25 °C)	4.00-4.02		4.02	

QC Release Date: 23.Mar.2017

The minimum shelf life is based on the current knowledge and holds only for proper storage conditions in the originally closed flasks/ packages.
We herewith confirm that the delivery is effected according to the technical delivery conditions agreed.
Particular properties of the products or the suitability for a particular area of application are not assured.
We guarantee a proper quality within our General Conditions of Sales.

Honeywell
Quality Management

This document was produced electronically and is valid without a signature