

**Honeywell****CERTIFICATE OF ANALYSIS****Honeywell Specialty Chemicals Seelze GmbH**

Wunstorferstrasse 40  
 Seelze, 30926  
 Telefon: +49 5137 999-0  
[www.lab-honeywell.com](http://www.lab-honeywell.com)

**Brand:** Honeywell Fluka™  
**Product:** 319481  
**Lot No:** J2180  
**Production date:** 06.Aug.2019  
**Expiry Date:** 29.Oct.2020

Sodium hydroxide solution volumetric, 0.1 M NaOH (0.1N)

Parameter	Specification	Units	Result	Units
assay titr.	0.095-0.105	mol/l	0.103	mol/l
carbonate (CO <sub>3</sub> )	max. 0.0015	mol/l	0.0014	mol/l
Traceable to NIST SRM	effektiv		60357	
BAM/EMPA ref.-material	effektiv		U2-0317	
BAM/EMPA Certificate no.	effektiv		SRM848	

**QC Release Date:** 20.Aug.2019

The minimum shelf life is based on the current knowledge and holds only for proper storage conditions in the originally closed flasks/ packages.

We herewith confirm that the delivery is effected according to the technical delivery conditions agreed.

Particular properties of the products or the suitability for a particular area of application are not assured.

We guarantee a proper quality within our General Conditions of Sales.

**Honeywell**  
**Quality Management**

**This document was produced electronically and is valid without a signature**